

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
 Data File : VN056821.D
 Acq On : 19 Jul 2019 11:02
 Operator : JC/SP
 Sample : PB121463ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121463ZHE#07

Quant Time: Jul 19 23:31:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	249139	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	450255	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	492838	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	173812	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	149707	54.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.64%	
35) Dibromofluoromethane	7.59	113	134346	47.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.90%	
50) Toluene-d8	10.09	98	578884	53.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.98%	
62) 4-Bromofluorobenzene	12.40	95	207409	56.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.66%	
Target Compounds						
16) Acetone	3.82	43	12396	9.761	ug/l	93
43) Isopropyl Acetate	8.17	43	17840	2.763	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
 Data File : VN056821.D
 Acq On : 19 Jul 2019 11:02
 Operator : JC/SP
 Sample : PB121463ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121463ZHE#07

Quant Time: Jul 19 23:31:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

