

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
 Data File : VN056830.D
 Acq On : 19 Jul 2019 14:31
 Operator : JC/SP
 Sample : PB121506TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121506TB

Quant Time: Jul 19 23:47:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	244016	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	450476	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	492462	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	173469	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	147113	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
35) Dibromofluoromethane	7.59	113	131819	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
50) Toluene-d8	10.09	98	574280	53.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.08%	
62) 4-Bromofluorobenzene	12.40	95	210443	57.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.26%	
Target Compounds						
16) Acetone	3.82	43	12728	10.233	ug/l	97
20) Methylene Chloride	4.56	84	3857	1.419	ug/l	85
43) Isopropyl Acetate	8.16	43	16722	2.589	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071919\
Data File : VN056830.D
Acq On : 19 Jul 2019 14:31
Operator : JC/SP
Sample : PB121506TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB121506TB

Quant Time: Jul 19 23:47:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

