

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
 Data File : VN056833.D
 Acq On : 19 Jul 2019 15:41
 Operator : JC/SP
 Sample : K3868-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Jul 20 00:05:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	260302	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	474006	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	512107	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	186286	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	151989	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
35) Dibromofluoromethane	7.59	113	138937	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
50) Toluene-d8	10.09	98	598558	52.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.08%	
62) 4-Bromofluorobenzene	12.40	95	221284	57.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.18%	
Target Compounds						
16) Acetone	3.82	43	16240	12.239	ug/l	97
18) Methyl Acetate	4.33	43	6850	1.351	ug/l #	88
20) Methylene Chloride	4.55	84	135943	46.883	ug/l	98
43) Isopropyl Acetate	8.17	43	14974	2.203	ug/l	93
95) Naphthalene	15.13	128	7578	5.216	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
Data File : VN056833.D
Acq On : 19 Jul 2019 15:41
Operator : JC/SP
Sample : K3868-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Jul 20 00:05:10 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

