

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
 Data File : VN056834.D
 Acq On : 19 Jul 2019 16:04
 Operator : JC/SP
 Sample : K2646-15DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-3DL

Quant Time: Jul 20 00:09:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	246643	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	443132	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	485080	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	187650	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	145776	53.43	ug/l	0.00
Spiked Amount			Recovery	=	106.86%	
35) Dibromofluoromethane	7.59	113	130984	47.50	ug/l	0.00
Spiked Amount			Recovery	=	95.00%	
50) Toluene-d8	10.09	98	565340	53.08	ug/l	0.00
Spiked Amount			Recovery	=	106.16%	
62) 4-Bromofluorobenzene	12.40	95	217180	60.46	ug/l	0.00
Spiked Amount			Recovery	=	120.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	113935	90.624	ug/l	100
17) Carbon Disulfide	4.05	76	4800	0.783	ug/l	98
18) Methyl Acetate	4.33	43	10625	2.845	ug/l #	89
20) Methylene Chloride	4.55	84	1759	0.640	ug/l	86
25) 2-Butanone	6.84	43	37099	20.184	ug/l	97
37) Ethyl Acetate	6.93	43	81485	20.162	ug/l	99
51) 4-Methyl-2-Pentanone	9.99	43	28551	6.938	ug/l	100
52) Toluene	10.16	92	56237	7.537	ug/l	97
68) m/p-Xylenes	11.62	106	1999	0.307	ug/l	84
84) 1,2,4-Trimethylbenzene	13.03	105	4130	0.330	ug/l	98
95) Naphthalene	15.13	128	9412	5.377	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071919\
Data File : VN056834.D
Acq On : 19 Jul 2019 16:04
Operator : JC/SP
Sample : K2646-15DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-3DL

Quant Time: Jul 20 00:09:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

