

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
 Data File : VN056837.D
 Acq On : 19 Jul 2019 17:14
 Operator : JC/SP
 Sample : K3615-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HR-06-071719-B

Quant Time: Jul 20 00:16:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	250808	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	451867	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	495140	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	173663	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	148152	53.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.80%	
35) Dibromofluoromethane	7.59	113	135392	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
50) Toluene-d8	10.09	98	581929	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
62) 4-Bromofluorobenzene	12.40	95	209327	57.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.30%	
Target Compounds						
16) Acetone	3.82	43	15241	11.921	ug/l	90
18) Methyl Acetate	4.34	43	7683	1.736	ug/l #	70
20) Methylene Chloride	4.54	84	20844661	7460.942	ug/l	99
43) Isopropyl Acetate	8.17	43	15796	2.438	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
Data File : VN056837.D
Acq On : 19 Jul 2019 17:14
Operator : JC/SP
Sample : K3615-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HR-06-071719-B

Quant Time: Jul 20 00:16:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

