

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
 Data File : VN056839.D
 Acq On : 19 Jul 2019 18:00
 Operator : JC/SP
 Sample : K3903-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 18107-NYCT-AVE-Z-COMP-3-4

Quant Time: Jul 20 00:21:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	249581	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	451523	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	495354	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	176070	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	146922	53.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.44%	
35) Dibromofluoromethane	7.59	113	133865	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
50) Toluene-d8	10.09	98	578982	53.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.70%	
62) 4-Bromofluorobenzene	12.40	95	213517	58.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.68%	
Target Compounds						
16) Acetone	3.82	43	19247	15.129	ug/l	99
18) Methyl Acetate	4.33	43	7229	1.587	ug/l #	71
20) Methylene Chloride	4.55	84	34783	12.511	ug/l	97
43) Isopropyl Acetate	8.17	43	15851	2.448	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
Data File : VN056839.D
Acq On : 19 Jul 2019 18:00
Operator : JC/SP
Sample : K3903-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
18107-NYCT-AVE-Z-COMP-3-4

Quant Time: Jul 20 00:21:41 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

