

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
 Data File : VN056840.D
 Acq On : 19 Jul 2019 18:24
 Operator : JC/SP
 Sample : K3868-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Jul 20 00:25:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	251014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	467005	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	518669	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	188921	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	153393	55.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.50%	
35) Dibromofluoromethane	7.59	113	139585	48.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.06%	
50) Toluene-d8	10.09	98	603089	53.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.46%	
62) 4-Bromofluorobenzene	12.40	95	223159	58.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.90%	
Target Compounds						
16) Acetone	3.82	43	15432	12.061	ug/l	93
18) Methyl Acetate	4.33	43	7840	1.790	ug/l	93
20) Methylene Chloride	4.55	84	24050	8.601	ug/l	98
30) Chloroform	7.38	83	10027	2.146	ug/l	98
43) Isopropyl Acetate	8.17	43	17267	2.579	ug/l #	86
95) Naphthalene	15.13	128	3079	4.802	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
Data File : VN056840.D
Acq On : 19 Jul 2019 18:24
Operator : JC/SP
Sample : K3868-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Jul 20 00:25:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

