

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
 Data File : VN056845.D
 Acq On : 19 Jul 2019 20:20
 Operator : JC/SP
 Sample : K3887-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AOC-7(2)

Quant Time: Jul 20 00:35:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	247643	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	449928	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	504264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	177411	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	147095	53.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.40%	
35) Dibromofluoromethane	7.59	113	135327	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
50) Toluene-d8	10.09	98	576131	53.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.56%	
62) 4-Bromofluorobenzene	12.40	95	214588	58.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.68%	
Target Compounds						
16) Acetone	3.82	43	15772	12.494	ug/l	99
18) Methyl Acetate	4.33	43	7810	1.817	ug/l	98
20) Methylene Chloride	4.55	84	5242	1.900	ug/l	95
43) Isopropyl Acetate	8.17	43	15614	2.420	ug/l #	90
95) Naphthalene	15.14	128	5112	5.015	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071919\
Data File : VN056845.D
Acq On : 19 Jul 2019 20:20
Operator : JC/SP
Sample : K3887-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AOC-7(2)

Quant Time: Jul 20 00:35:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

