


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN071919\
 Data File : VN056818.D
 Acq On : 19 Jul 2019 9:33
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Title : SW846 8260
 Last Update : Tue Jul 16 07:03:34 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.6	7446	PASS
75	95	30	60	44.3	21122	PASS
95	95	100	100	100.0	47642	PASS
96	95	5	9	6.4	3052	PASS
173	174	0.00	2	0.9	405	PASS
174	95	50	100	90.0	42872	PASS
175	174	5	9	7.7	3292	PASS
176	174	95	101	97.8	41930	PASS
177	176	5	9	5.9	2468	PASS