

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067790.D
 Acq On : 19 Jul 2021 13:50
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 19 18:55:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	393542	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	729877	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	689854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	257917	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	313571	52.722	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.440%
35) Dibromofluoromethane	8.024	113	227608	51.615	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.220%
50) Toluene-d8	10.443	98	951432	51.546	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.100%
62) 4-Bromofluorobenzene	12.733	95	315999	48.463	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
Data File : VN067790.D
Acq On : 19 Jul 2021 13:50
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 19 18:55:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

