

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067793.D
 Acq On : 19 Jul 2021 15:13
 Operator : JC/MD
 Sample : M2968-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jul 19 18:57:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	418603	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	772142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	724110	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	264999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	311668	49.265	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.520%
35) Dibromofluoromethane	8.024	113	228570	48.996	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.000%
50) Toluene-d8	10.443	98	952011	48.754	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.500%
62) 4-Bromofluorobenzene	12.736	95	310625	45.031	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.060%
Target Compounds						
						Qvalue
16) Acetone	4.304	43	7712	3.874	ug/l	99
95) Naphthalene	15.507	128	9925	3.731	ug/l	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
Data File : VN067793.D
Acq On : 19 Jul 2021 15:13
Operator : JC/MD
Sample : M2968-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Jul 19 18:57:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

