

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067794.D
 Acq On : 19 Jul 2021 15:37
 Operator : JC/MD
 Sample : PB137761TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137761TB

Quant Time: Jul 20 04:52:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	319667	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	589946	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	546425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	196586	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	247543	51.239	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.480%
35) Dibromofluoromethane	8.024	113	178233	50.005	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.000%
50) Toluene-d8	10.446	98	742397	49.761	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.520%
62) 4-Bromofluorobenzene	12.733	95	241281	45.781	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.560%
Target Compounds						
20) Methylene Chloride	5.095	84	16958	4.140	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
Data File : VN067794.D
Acq On : 19 Jul 2021 15:37
Operator : JC/MD
Sample : PB137761TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137761TB

Quant Time: Jul 20 04:52:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

