

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067797.D
 Acq On : 19 Jul 2021 16:47
 Operator : JC/MD
 Sample : PB137749ZHE#19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137749ZHE#19

Manual Integrations
 APPROVED

MMDadoda
 7/20/2021 5:39:38 PM

Quant Time: Jul 20 04:53:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	329185	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	619736	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	645794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	232709	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	257289	51.716	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.440%			
35) Dibromofluoromethane	8.024	113	187529	50.084	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.160%			
50) Toluene-d8	10.446	98	806816	51.479	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.960%			
62) 4-Bromofluorobenzene	12.733	95	275832	49.821	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.640%			
Target Compounds					Qvalue	
16) Acetone	4.304	43	22238m	14.206	ug/l	
20) Methylene Chloride	5.093	84	23906	5.928	ug/l	87
43) Isopropyl Acetate	8.566	43	40612	4.139	ug/l #	93
95) Naphthalene	15.507	128	16324	4.359	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
Data File : VN067797.D
Acq On : 19 Jul 2021 16:47
Operator : JC/MD
Sample : PB137749ZHE#19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137749ZHE#19

Manual Integrations
APPROVED

MMDadoda
7/20/2021 5:39:38 PM

Quant Time: Jul 20 04:53:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

