

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067798.D
 Acq On : 19 Jul 2021 17:12
 Operator : JC/MD
 Sample : M3061-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB-2021-WS-000-08

Quant Time: Jul 20 04:53:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	313725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	582183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	606887	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	224449	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	243101	51.273	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.540%			
35) Dibromofluoromethane	8.024	113	180098	51.202	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.400%			
50) Toluene-d8	10.446	98	771901	52.428	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.860%			
62) 4-Bromofluorobenzene	12.733	95	265896	51.124	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.240%			
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	65764	44.080	ug/l	97
20) Methylene Chloride	5.092	84	25089	6.600	ug/l	91
25) 2-Butanone	7.351	43	19617	7.918	ug/l	96
43) Isopropyl Acetate	8.563	43	37131	4.028	ug/l #	92
44) Trichloroethene	9.220	130	24159	5.881	ug/l	90
64) Tetrachloroethene	10.979	164	12127	2.831	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
Data File : VN067798.D
Acq On : 19 Jul 2021 17:12
Operator : JC/MD
Sample : M3061-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB-2021-WS-000-08

Quant Time: Jul 20 04:53:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

