

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067799.D
 Acq On : 19 Jul 2021 17:36
 Operator : JC/MD
 Sample : M3088-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B1-A-TOP-(0)

Quant Time: Jul 20 04:53:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	314214	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	585842	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	611991	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	223291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	245327	51.662	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	103.320%	
35) Dibromofluoromethane	8.024	113	181027	51.145	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	102.280%	
50) Toluene-d8	10.443	98	768469	51.869	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	103.740%	
62) 4-Bromofluorobenzene	12.734	95	266277	50.878	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	101.760%	
Target Compounds						
						Qvalue
16) Acetone	4.296	43	35664	23.868	ug/l	# 88
20) Methylene Chloride	5.087	84	28585	7.605	ug/l	89
43) Isopropyl Acetate	8.563	43	38694	4.172	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067799.D
 Acq On : 19 Jul 2021 17:36
 Operator : JC/MD
 Sample : M3088-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleID :
 B1-A-TOP-(0)

Quant Time: Jul 20 04:53:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

