

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067803.D
 Acq On : 19 Jul 2021 19:15
 Operator : JC/MD
 Sample : M3088-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B1-C-TOP-(0)

Manual Integrations
 APPROVED

MMDadoda
 7/20/2021 5:39:41 PM

Quant Time: Jul 20 04:54:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	320355	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	608993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	633254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	233228	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	254031	52.469	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	104.940%	
35) Dibromofluoromethane	8.027	113	182200	49.519	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	99.040%	
50) Toluene-d8	10.443	98	796486	51.717	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	103.440%	
62) 4-Bromofluorobenzene	12.734	95	272120	50.018	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	100.040%	
Target Compounds						Qvalue
16) Acetone	4.296	43	37680	24.734	ug/l	95
18) Methyl Acetate	4.857	43	8972m	2.097	ug/l	
20) Methylene Chloride	5.101	84	29536	7.717	ug/l	86
43) Isopropyl Acetate	8.566	43	35708	3.704	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
Data File : VN067803.D
Acq On : 19 Jul 2021 19:15
Operator : JC/MD
Sample : M3088-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B1-C-TOP-(0)

Manual Integrations
APPROVED

MMDadoda
7/20/2021 5:39:41 PM

Quant Time: Jul 20 04:54:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

