

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067804.D
 Acq On : 19 Jul 2021 19:39
 Operator : JC/MD
 Sample : M3088-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B1-D-TOP-(0)

Quant Time: Jul 20 04:54:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	315940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	606020	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	608274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	227780	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	251405	52.652	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 105.300%		
35) Dibromofluoromethane	8.024	113	183202	50.036	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 100.080%		
50) Toluene-d8	10.446	98	784163	51.166	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 102.340%		
62) 4-Bromofluorobenzene	12.734	95	268496	49.594	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 99.180%		
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	54048	35.973	ug/l	92
18) Methyl Acetate	4.873	43	7150	1.694	ug/l #	67
20) Methylene Chloride	5.095	84	28753	7.608	ug/l	87
43) Isopropyl Acetate	8.566	43	31219	3.254	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
Data File : VN067804.D
Acq On : 19 Jul 2021 19:39
Operator : JC/MD
Sample : M3088-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B1-D-TOP-(0)

Quant Time: Jul 20 04:54:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

