

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067805.D
 Acq On : 19 Jul 2021 20:03
 Operator : JC/MD
 Sample : M3088-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B1-D-BOTTOM-(4-2)

Quant Time: Jul 20 04:54:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	316874	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	596628	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	616258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	231939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	250163	52.238	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.480%	
35) Dibromofluoromethane	8.024	113	180092	49.961	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.920%	
50) Toluene-d8	10.443	98	781940	51.824	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.640%	
62) 4-Bromofluorobenzene	12.733	95	269135	50.494	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.980%	
Target Compounds						
						Qvalue
16) Acetone	4.293	43	67827	45.011	ug/l	91
18) Methyl Acetate	4.867	43	7441	1.758	ug/l #	65
20) Methylene Chloride	5.093	84	32274	8.598	ug/l	90
25) 2-Butanone	7.351	43	16802	6.715	ug/l	91
43) Isopropyl Acetate	8.566	43	32423	3.433	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
Data File : VN067805.D
Acq On : 19 Jul 2021 20:03
Operator : JC/MD
Sample : M3088-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B1-D-BOTTOM-(4-2)

Quant Time: Jul 20 04:54:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

