

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
 Data File : VN067807.D
 Acq On : 19 Jul 2021 20:53
 Operator : JC/MD
 Sample : M3089-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-N2-A-BOTTOM-(12)

Manual Integrations
 APPROVED

MMDadoda
 7/20/2021 5:39:43 PM

Quant Time: Jul 20 04:55:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	302553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	580273	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	601934	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	225628	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	241069	52.721	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 105.440%			
35) Dibromofluoromethane	8.024	113	176966	50.477	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.960%			
50) Toluene-d8	10.443	98	762862	51.985	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.960%			
62) 4-Bromofluorobenzene	12.733	95	261789	50.500	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 101.000%			
Target Compounds					Qvalue	
16) Acetone	4.293	43	37148	25.819	ug/l	98
18) Methyl Acetate	4.856	43	8275m	2.048	ug/l	
20) Methylene Chloride	5.098	84	33089	9.285	ug/l	88
43) Isopropyl Acetate	8.563	43	32754	3.565	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071921\
Data File : VN067807.D
Acq On : 19 Jul 2021 20:53
Operator : JC/MD
Sample : M3089-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-N2-A-BOTTOM-(12)

Manual Integrations
APPROVED

MMDadoda
7/20/2021 5:39:43 PM

Quant Time: Jul 20 04:55:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

