

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071922\
 Data File : VN073513.D
 Acq On : 19 Jul 2022 13:50
 Operator : JC\MD
 Sample : VN0719WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0719WBL01

Quant Time: Jul 20 06:54:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	252086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	428400	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	395375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	189020	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	173075	47.559	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.120%
35) Dibromofluoromethane	7.951	113	146085	51.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.500%
50) Toluene-d8	10.380	98	469558	46.385	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.780%
62) 4-Bromofluorobenzene	12.669	95	193588	49.763	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071922\
Data File : VN073513.D
Acq On : 19 Jul 2022 13:50
Operator : JC\MD
Sample : VN0719WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0719WBL01

Quant Time: Jul 20 06:54:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

