

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071922\
 Data File : VN073516.D
 Acq On : 19 Jul 2022 15:15
 Operator : JC\MD
 Sample : N3765-01 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPC-91BN-84

Quant Time: Jul 20 06:25:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	243605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	416242	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	388575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	182570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	160319	45.588	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.180%		
35) Dibromofluoromethane	7.951	113	135583	48.959	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.920%		
50) Toluene-d8	10.381	98	444114	45.153	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.300%		
62) 4-Bromofluorobenzene	12.669	95	182484	48.279	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.560%		
Target Compounds						Qvalue
20) Methylene Chloride	5.004	84	2434	0.702	ug/l #	91
30) Chloroform	7.751	83	4861	0.827	ug/l	98
38) Carbon Tetrachloride	8.128	117	7366	1.672	ug/l	86
44) Trichloroethene	9.151	130	1218	0.364	ug/l	50
64) Tetrachloroethene	10.916	164	3270093	1030.600	ug/l	91
90) Hexachloroethane	14.151	117	34176	14.675	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071922\
Data File : VN073516.D
Acq On : 19 Jul 2022 15:15
Operator : JC\MD
Sample : N3765-01 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPC-91BN-84

Quant Time: Jul 20 06:25:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

