

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071922\
 Data File : VN073518.D
 Acq On : 19 Jul 2022 16:04
 Operator : JC\MD
 Sample : N3765-02 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPC-01F-84

Quant Time: Jul 20 06:25:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	238772	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	411450	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	395408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	186776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	163543	47.446	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.900%
35) Dibromofluoromethane	7.951	113	135808	49.611	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.220%
50) Toluene-d8	10.380	98	447266	46.003	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.000%
62) 4-Bromofluorobenzene	12.674	95	189301	50.665	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.340%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.163	130	976	0.295	ug/l	58
64) Tetrachloroethene	10.916	164	1037735	321.400	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071922\
Data File : VN073518.D
Acq On : 19 Jul 2022 16:04
Operator : JC\MD
Sample : N3765-02 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPC-01F-84

Quant Time: Jul 20 06:25:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

