

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071922\
 Data File : VN073521.D
 Acq On : 19 Jul 2022 21:33
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 20 06:25:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	211162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	368906	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	351659	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	159275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	151170	49.591	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.180%
35) Dibromofluoromethane	7.951	113	129198	52.640	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.280%
50) Toluene-d8	10.380	98	399467	45.825	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.660%
62) 4-Bromofluorobenzene	12.674	95	160445	47.895	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.780%
Target Compounds						
					Qvalue	
64) Tetrachloroethene	10.916	164	14417	5.021	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071922\
Data File : VN073521.D
Acq On : 19 Jul 2022 21:33
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 20 06:25:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

