

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071923\
 Data File : VN078576.D
 Acq On : 19 Jul 2023 13:24
 Operator : JC\MD
 Sample : 03664-03 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1007

Quant Time: Jul 20 01:32:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	396021	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	709978	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	649767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	242364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	258818	48.750	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.500%
35) Dibromofluoromethane	8.171	113	262977	54.565	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.120%
50) Toluene-d8	10.571	98	835976	45.916	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.840%
62) 4-Bromofluorobenzene	12.859	95	282959	50.456	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.920%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.488	105	39710	2.272	ug/l	97
95) Naphthalene	15.647	128	20783	2.189	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071923\
Data File : VN078576.D
Acq On : 19 Jul 2023 13:24
Operator : JC\MD
Sample : 03664-03 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1007

Quant Time: Jul 20 01:32:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

