

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071923\
 Data File : VN078579.D
 Acq On : 19 Jul 2023 14:36
 Operator : JC\MD
 Sample : 03664-06 200X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1022

Quant Time: Jul 20 01:33:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	406704	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	753129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	701815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	276687	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	274903	50.420	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.840%	
35) Dibromofluoromethane	8.177	113	278208	54.417	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.840%	
50) Toluene-d8	10.571	98	871714	45.135	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 90.280%	
62) 4-Bromofluorobenzene	12.853	95	319680	53.737	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.480%	
Target Compounds						Qvalue
16) Acetone	4.447	43	15066	8.871	ug/l	97
84) 1,2,4-Trimethylbenzene	13.494	105	5441	0.273	ug/l	94
95) Naphthalene	15.641	128	14332	1.322	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071923\
Data File : VN078579.D
Acq On : 19 Jul 2023 14:36
Operator : JC\MD
Sample : 03664-06 200X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1022

Quant Time: Jul 20 01:33:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

