

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071924\
 Data File : VN082987.D
 Acq On : 19 Jul 2024 14:34
 Operator : JC\MD
 Sample : P3307-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VPB189-HYD-20240717

Quant Time: Jul 22 02:40:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	69448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	136472	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	137954	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	53827	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	55757	57.274	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.540%
35) Dibromofluoromethane	8.171	113	46857	53.508	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.020%
50) Toluene-d8	10.571	98	160946	49.761	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.520%
62) 4-Bromofluorobenzene	12.847	95	48563	41.799	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	83.600%
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	1409	4.111	ug/l	98
60) Dibromochloromethane	11.359	129	482	0.444	ug/l	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071924\
Data File : VN082987.D
Acq On : 19 Jul 2024 14:34
Operator : JC\MD
Sample : P3307-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VPB189-HYD-20240717

Quant Time: Jul 22 02:40:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

