

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071924\
 Data File : VN082990.D
 Acq On : 19 Jul 2024 15:46
 Operator : JC\MD
 Sample : P3307-01RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-189-TB-20240716RE

Quant Time: Jul 22 02:06:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	67994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	127288	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	111196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	44824	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	55479	58.207	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.420%
35) Dibromofluoromethane	8.165	113	42702	52.281	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.560%
50) Toluene-d8	10.571	98	141563	46.926	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.860%
62) 4-Bromofluorobenzene	12.853	95	40896	37.739	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	75.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071924\
Data File : VN082990.D
Acq On : 19 Jul 2024 15:46
Operator : JC\MD
Sample : P3307-01RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-189-TB-20240716RE

Quant Time: Jul 22 02:06:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

