

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071924\
 Data File : VN082992.D
 Acq On : 19 Jul 2024 16:35
 Operator : JC\MD
 Sample : P3307-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VPB189-HYD-20240717

Quant Time: Jul 22 02:07:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	73464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	143735	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	142172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	57444	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	57162	55.507	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.020%			
35) Dibromofluoromethane	8.171	113	46950	50.905	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.800%			
50) Toluene-d8	10.571	98	162112	47.589	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.180%			
62) 4-Bromofluorobenzene	12.847	95	53205	43.480	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 86.960%			
Target Compounds						
16) Acetone	4.436	43	1889	5.210	ug/l	92
60) Dibromochloromethane	11.359	129	562	0.491	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071924\
Data File : VN082992.D
Acq On : 19 Jul 2024 16:35
Operator : JC\MD
Sample : P3307-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VPB189-HYD-20240717

Quant Time: Jul 22 02:07:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

