

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
 Data File : VN049998.D
 Acq On : 21 Jul 2018 8:42
 Operator : MD\SY
 Sample : J3820-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BU-01-071918

Quant Time: Jul 23 01:05:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	577006	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	972390	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	797160	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	283844	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	418363	52.75	ug/l	0.00
Spiked Amount			Recovery	=	105.50%	
35) Dibromofluoromethane	7.59	113	366752	44.93	ug/l	0.00
Spiked Amount			Recovery	=	89.86%	
50) Toluene-d8	10.09	98	1295742	43.56	ug/l	0.00
Spiked Amount			Recovery	=	87.12%	
62) 4-Bromofluorobenzene	12.40	95	342339	33.42	ug/l	0.00
Spiked Amount			Recovery	=	66.84%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.57	94	361	Below Cal	#	45
16) Acetone	3.83	43	29748	14.49	ug/l	100
18) Methyl Acetate	4.33	43	25015	3.10	ug/l	96
20) Methylene Chloride	4.55	84	52189	6.68	ug/l	98
25) 2-Butanone	6.85	43	12622	4.30	ug/l	99
29) Tetrahydrofuran	7.22	42	3430	1.86	ug/l	# 57
30) Chloroform	7.38	83	5946	0.45	ug/l	96
31) Cyclohexane	7.66	56	10800	1.02	ug/l	# 4
36) 1,1-Dichloropropene	7.66	75	41901	3.88	ug/l	# 49
37) Ethyl Acetate	6.85	43	12622	1.70	ug/l	# 68
38) Carbon Tetrachloride	7.67	117	53057	4.58	ug/l	# 16
41) Methacrylonitrile	7.22	41	4475	1.20	ug/l	# 36
43) Isopropyl Acetate	8.17	43	256703	19.79	ug/l	95
48) Methyl methacrylate	9.18	41	7891	1.21	ug/l	# 20
51) 4-Methyl-2-Pentanone	9.99	43	1768	0.24	ug/l	92
56) Ethyl methacrylate	10.23	69	2136	2.34	ug/l	# 75
59) 2-Hexanone	10.77	43	15993	3.28	ug/l	# 61
68) m/p-Xylenes	11.62	106	3107	0.25	ug/l	88
71) Bromoform	12.40	173	1530	0.28	ug/l	# 28
76) 1,2,3-Trichloropropane	12.40	75	170076	31.94	ug/l	# 100
81) trans-1,4-Dichloro-2-buten	12.40	75	170076	99.88	ug/l	# 13
84) 1,2,4-Trimethylbenzene	13.04	105	4805	0.27	ug/l	97
90) Hexachloroethane	13.89	117	1364	0.36	ug/l	# 4
93) 1,2,4-Trichlorobenzene	14.92	180	61	3.49	ug/l	# 1
95) Naphthalene	15.14	128	23488	6.08	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072018\
Data File : VN049998.D
Acq On : 21 Jul 2018 8:42
Operator : MD\SY
Sample : J3820-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BU-01-071918

Quant Time: Jul 23 01:05:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. 0.00 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampled :


BU-01-071918

Tot Ion:168 Resp: 577006

Ion Ratio Lower Upper




168 100

99 53.3 40.6 61.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#5

Bromomethane

Concen: Below Cal

RT: 2.57 min Scan# 304

Delta R.T. 0.01 min

Lab File: VN049998.D


Acq: 21 Jul 2018 8:42

Tgt Ion: 94 Resp: 361

Ion Ratio Lower Upper



94 100


96 41.2 75.7 113.5#



Abundance Ion 93.90 (93.60 to 94.60): VNO

Ion 95.90 (95.60 to 96.60): VNO





#16

Acetone

Concen: 14.49 ug/l

RT: 3.83 min Scan# 696

Delta R.T. 0.01 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampleId :


BU-01-071918

Tot Ion: 43 Resp: 29748

Ion Ratio Lower Upper




43 100

58 32.5 25.9 38.9



Abundance Ion 43.00 (42.70 to 43.70): VN049998.D (-601) (-)

Ion 58.00 (57.70 to 58.70): VN049998.D (-601) (-)



#18

Methyl Acetate

Concen: 3.10 ug/l

RT: 4.33 min Scan# 853

Delta R.T. 0.00 min

Lab File: VN049998.D


Acq: 21 Jul 2018 8:42

Tgt Ion: 43 Resp: 25015

Ion Ratio Lower Upper



43 100


74 22.6 19.8 29.8



Abundance Ion 43.00 (42.70 to 43.70): VN049998.D (-759) (-)

Ion 74.00 (73.70 to 74.70): VN049998.D (-759) (-)









#20
Methvlene Chloride
Concen: 6.68 ug/l
RT: 4.55 min Scan# 921
Delta R.T. 0.00 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42




Instrument :
MSVOA_N
ClientSampleId :
BU-01-071918


Tot Ion: 84 Resp: 52189
Ion Ratio Lower Upper
84 100
49 120.8 95.4 143.2
51 36.0 28.4 42.6
86 62.4 52.6 79.0



#25
2-Butanone
Concen: 4.30 ug/l
RT: 6.85 min Scan# 1637
Delta R.T. 0.01 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42

Tgt Ion: 43 Resp: 12622
Ion Ratio Lower Upper
43 100
72 28.2 22.1 33.1





#29

Tetrahydrofuran

Concen: 1.86 ug/l

RT: 7.22 min Scan# 1750

Delta R.T. 0.00 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampled :

BU-01-071918


Tot Ion: 42 Resp: 3430

Ion Ratio Lower Upper

42 100

72 61.1 35.6 53.4#




71 81.0 34.0 51.0#



Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0



#30

Chloroform

Concen: 0.45 ug/l

RT: 7.38 min Scan# 1800

Delta R.T. 0.00 min

Lab File: VN049998.D


Acq: 21 Jul 2018 8:42

Tgt Ion: 83 Resp: 5946

Ion Ratio Lower Upper



83 100


85 61.2 51.4 77.2



Abundance Ion 82.90 (82.60 to 83.60): VN0

Ion 84.90 (84.60 to 85.60): VN0





#31

Cyclohexane

Concen: 1.02 ug/l

RT: 7.66 min Scan# 1889

Delta R.T. 0.01 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampled :

BU-01-071918


Tot Ion: 56 Resp: 10800

Ion Ratio Lower Upper

56 100

69 168.9 26.6 40.0#




84 123.9 69.0 103.6#



Abundance Ion 56.00 (55.70 to 56.70): VN049866.D (-1863) (-)

Ion 69.00 (68.70 to 69.70): VN049866.D (-1863) (-)

Ion 84.00 (83.70 to 84.70): VN049866.D (-1863) (-)



#33

1,2-Dichloroethane-d4

Concen: 52.75 ug/l

RT: 8.03 min Scan# 2003

Delta R.T. 0.00 min

Lab File: VN049998.D


Acq: 21 Jul 2018 8:42

Tgt Ion: 65 Resp: 418363

Ion Ratio Lower Upper



65 100


67 52.6 0.0 107.2



Abundance Ion 64.90 (64.60 to 65.60): VN049866.D (-1983) (-)

Ion 66.90 (66.60 to 67.60): VN049866.D (-1983) (-)





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 8.59 min Scan# 2176

Delta R.T. 0.00 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampleId :

BU-01-071918


Tot Ion:114 Resp: 972390

Ion Ratio Lower Upper

114 100

63 20.3 0.0 39.8




88 15.3 0.0 30.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 44.93 ug/l

RT: 7.59 min Scan# 1867

Delta R.T. 0.00 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42


Tgt Ion:113 Resp: 366752

Ion Ratio Lower Upper

113 100

111 102.3 81.8 122.8



192 22.2 17.8 26.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 3.88 ug/l
 RT: 7.66 min Scan# 1889
 Delta R.T. -0.13 min
 Lab File: VN049998.D
 Acq: 21 Jul 2018 8:42




Instrument :
 MSVOA_N
 ClientSampleId :
 BU-01-071918


Tot Ion	75	Resp	41901
Ion Ratio	Lower	Upper	
75	100		
110	8.5	18.1	54.4#
77	0.0	24.9	37.3#



#37
 Ethyl Acetate
 Concen: 1.70 ug/l
 RT: 6.85 min Scan# 1637
 Delta R.T. -0.08 min
 Lab File: VN049998.D
 Acq: 21 Jul 2018 8:42

Tgt Ion	43	Resp	12622
Ion Ratio	Lower	Upper	
43	100		
61	0.0	11.7	17.5#
70	0.0	8.4	12.6#









#38
Carbon Tetrachloride
Concen: 4.58 ug/l
RT: 7.67 min Scan# 1890
Delta R.T. -0.11 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42

Instrument :
MSVOA_N
ClientSampled :
BU-01-071918

Tot Ion:117 Resp: 53057
Ion Ratio Lower Upper
117 100
119 5.8 77.2 115.8#
121 0.0 25.0 37.6#




Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V






#41
Methacrylonitrile
Concen: 1.20 ug/l
RT: 7.22 min Scan# 1750
Delta R.T. 0.04 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42

Tgt Ion: 41 Resp: 4475
Ion Ratio Lower Upper
41 100
39 36.8 45.4 68.2#
67 0.0 65.4 98.0#
52 0.0 24.7 37.1#



Abundance Ion 41.00 (40.70 to 41.70): VNO
Ion 39.00 (38.70 to 39.70): VNO
Ion 67.00 (66.70 to 67.70): VNO
Ion 52.00 (51.70 to 52.70): VNO





#43

Isopropyl Acetate

Concen: 19.79 ug/l

RT: 8.17 min Scan# 2047

Delta R.T. 0.00 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampled :

BU-01-071918


Tot Ion: 43 Resp: 256703

Ion Ratio Lower Upper




43 100

61 16.5 14.9 22.3

87 10.2 10.0 15.0



Abundance Ion 43.00 (42.70 to 43.70): VNO
Ion 61.00 (60.70 to 61.70): VNO
Ion 87.00 (86.70 to 87.70): VNO



#48

Methyl methacrylate

Concen: 1.21 ug/l

RT: 9.18 min Scan# 2361

Delta R.T. -0.02 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42


Tgt Ion: 41 Resp: 7891

Ion Ratio Lower Upper



41 100


69 0.0 71.5 107.3#

39 16.2 44.7 67.1#



Abundance Ion 41.00 (40.70 to 41.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 39.00 (38.70 to 39.70): VNO





#50

Toluene-d8

Concen: 43.56 ug/l

RT: 10.09 min Scan# 2644

Delta R.T. 0.00 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampleId :





BU-01-071918

Tot Ion: 98 Resp: 1295742

Ion Ratio Lower Upper

98 100

100 63.1 51.7 77.5



#51

4-Methyl-2-Pentanone

Concen: 0.24 ug/l

RT: 9.99 min Scan# 2611

Delta R.T. 0.00 min

Lab File: VN049998.D




Acq: 21 Jul 2018 8:42


Tgt Ion: 43 Resp: 1768

Ion Ratio Lower Upper

43 100

58 35.1 31.8 47.8









#56
Ethyl methacrylate
Concen: 2.34 ug/l
RT: 10.23 min Scan# 2688
Delta R.T. -0.20 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42




Instrument :
MSVOA_N
ClientSampled :
BU-01-071918


Tot Ion: 69 Resp: 2136
Ion Ratio Lower Upper
69 100
41 43.4 51.6 77.4#
39 21.0 25.8 38.8#



#59
2-Hexanone
Concen: 3.28 ug/l
RT: 10.77 min Scan# 2854
Delta R.T. 0.01 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42

Tgt Ion: 43 Resp: 15993
Ion Ratio Lower Upper
43 100
58 26.9 27.8 83.4#









#62
4-Bromofluorobenzene
Concen: 33.42 ug/l
RT: 12.40 min Scan# 3363
Delta R.T. 0.00 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42




Instrument :
MSVOA_N
ClientSampleId :
BU-01-071918


Tot Ion	95	Resp	342339
Ion Ratio	Lower	Upper	
95	100		
174	92.2	0.0	179.6
176	87.0	0.0	174.0



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. 0.00 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42

Tgt Ion	117	Resp	797160
Ion Ratio	Lower	Upper	
117	100		
82	54.1	41.9	62.9
119	32.3	25.4	38.2









#68
m/p-Xvlenes
Concen: 0.25 ug/l
RT: 11.62 min Scan# 3119
Delta R.T. -0.00 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42




Instrument :
MSVOA_N
ClientSampleId :
BU-01-071918


Tot Ion:106 Resp: 3107
Ion Ratio Lower Upper
106 100
91 185.2 163.4 245.2



#71
Bromoform
Concen: 0.28 ug/l
RT: 12.40 min Scan# 3362
Delta R.T. 0.27 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42

Tgt Ion:173 Resp: 1530
Ion Ratio Lower Upper
173 100
175 0.0 24.6 73.8#
254 0.0 0.2 0.2#





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. 0.00 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampleId :

BU-01-071918


Tot Ion:152 Resp: 283844

Ion Ratio Lower Upper

152 100

115 56.0 28.1 84.2




150 155.2 0.0 347.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 31.94 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN049998.D


Acq: 21 Jul 2018 8:42

Tgt Ion: 75 Resp: 170076

Ion Ratio Lower Upper



75 100


77 1.3 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): V

Ion 77.00 (76.70 to 77.70): V





#81

trans-1,4-Dichloro-2-butene

Concen: 99.88 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampleId :

BU-01-071918


Tot Ion: 75 Resp: 170076

Ion Ratio Lower Upper

75 100

53 0.2 73.9 110.9#




89 0.0 35.9 53.9#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0



#84

1,2,4-Trimethylbenzene

Concen: 0.27 ug/l

RT: 13.04 min Scan# 3562

Delta R.T. 0.00 min

Lab File: VN049998.D


Acq: 21 Jul 2018 8:42

Tgt Ion: 105 Resp: 4805

Ion Ratio Lower Upper



105 100


120 43.6 22.7 68.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#90

Hexachloroethane

Concen: 0.36 ug/l

RT: 13.89 min Scan# 3826

Delta R.T. 0.02 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42

Instrument :

MSVOA_N

ClientSampled :


BU-01-071918

Tot Ion:117 Resp: 1364

Ion Ratio Lower Upper




117 100

201 0.0 45.3 135.8#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#93

1,2,4-Trichlorobenzene

Concen: 3.49 ug/l

RT: 14.92 min Scan# 4146

Delta R.T. 0.01 min

Lab File: VN049998.D

Acq: 21 Jul 2018 8:42


Tgt Ion:180 Resp: 61

Ion Ratio Lower Upper

180 100

182 0.0 47.3 141.9#



145 708.2 14.4 43.0#




Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.80 (144.50 to 145.50): V






#95
Naphthalene
Concen: 6.08 ug/l
RT: 15.14 min Scan# 4213
Delta R.T. 0.00 min
Lab File: VN049998.D
Acq: 21 Jul 2018 8:42

Instrument :
MSVOA_N
ClientSampleId :
BU-01-071918

Tot Ion	Ratio	Lower	Upper
128	100		
127	12.8	10.4	15.6
129	11.2	8.6	13.0



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V

