

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
 Data File : VN049969.D
 Acq On : 20 Jul 2018 18:05
 Operator : MD\SY
 Sample : PB111289ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111289ZHE#09

Quant Time: Jul 21 02:19:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	546353	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	938165	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	772653	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	270712	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	402023	53.54	ug/l	0.00
Spiked Amount						
			Recovery	=	107.08%	
35) Dibromofluoromethane	7.59	113	352814	44.80	ug/l	0.00
Spiked Amount						
			Recovery	=	89.60%	
50) Toluene-d8	10.09	98	1239673	43.20	ug/l	0.00
Spiked Amount						
			Recovery	=	86.40%	
62) 4-Bromofluorobenzene	12.40	95	328434	33.23	ug/l	0.00
Spiked Amount						
			Recovery	=	66.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	28603	14.71	ug/l	90
18) Methyl Acetate	4.34	43	28922	3.97	ug/l	100
20) Methylene Chloride	4.55	84	15359	2.08	ug/l	96
25) 2-Butanone	6.85	43	46598	16.77	ug/l	92
29) Tetrahydrofuran	7.22	42	23356	13.34	ug/l	97
43) Isopropyl Acetate	8.17	43	317659	25.38	ug/l #	92
59) 2-Hexanone	10.77	43	19855	4.23	ug/l #	59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072018\
Data File : VN049969.D
Acq On : 20 Jul 2018 18:05
Operator : MD\SY
Sample : PB111289ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111289ZHE#09

Quant Time: Jul 21 02:19:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

