

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
 Data File : VN049973.D
 Acq On : 20 Jul 2018 19:49
 Operator : MD\SY
 Sample : J4016-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB5-U-23-13

Quant Time: Jul 21 02:29:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	549982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	936344	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	793052	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	300309	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	405485	53.64	ug/l	0.00
Spiked Amount						
			Recovery	=	107.28%	
35) Dibromofluoromethane	7.59	113	357778	45.51	ug/l	0.00
Spiked Amount						
			Recovery	=	91.02%	
50) Toluene-d8	10.09	98	1247396	43.55	ug/l	0.00
Spiked Amount						
			Recovery	=	87.10%	
62) 4-Bromofluorobenzene	12.40	95	344222	34.90	ug/l	0.00
Spiked Amount						
			Recovery	=	69.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	25336	12.95	ug/l	98
18) Methyl Acetate	4.34	43	28448	3.86	ug/l	98
20) Methylene Chloride	4.55	84	309247	41.55	ug/l	99
43) Isopropyl Acetate	8.17	43	244549	19.58	ug/l	96
59) 2-Hexanone	10.77	43	17602	3.75	ug/l #	58
95) Naphthalene	15.14	128	59716	8.31	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072018\
Data File : VN049973.D
Acq On : 20 Jul 2018 19:49
Operator : MD\SY
Sample : J4016-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB5-U-23-13

Quant Time: Jul 21 02:29:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

