

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
 Data File : VN049979.D
 Acq On : 20 Jul 2018 22:25
 Operator : MD\SY
 Sample : J4057-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LS-SOIL-3

Quant Time: Jul 21 03:20:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	528909	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	929321	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	773154	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	284972	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	393424	54.12	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	108.24%	
35) Dibromofluoromethane	7.59	113	343168	43.99	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	87.98%	
50) Toluene-d8	10.09	98	1230927	43.30	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	86.60%	
62) 4-Bromofluorobenzene	12.40	95	339818	34.71	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	69.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	13232	7.03	ug/l #	88
18) Methyl Acetate	4.34	43	28302	4.03	ug/l	95
20) Methylene Chloride	4.55	84	99031	13.84	ug/l	97
43) Isopropyl Acetate	8.17	43	289584	23.36	ug/l #	93
59) 2-Hexanone	10.77	43	17687	3.80	ug/l #	57
95) Naphthalene	15.14	128	9105	5.11	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072018\
Data File : VN049979.D
Acq On : 20 Jul 2018 22:25
Operator : MD\SY
Sample : J4057-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LS-SOIL-3

Quant Time: Jul 21 03:20:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

