

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
 Data File : VN049980.D
 Acq On : 20 Jul 2018 22:51
 Operator : MD\SY
 Sample : J4057-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LS-SOIL-2

Quant Time: Jul 21 03:22:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	544564	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	946069	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	793274	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	291787	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	404618	54.06	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	108.12%	
35) Dibromofluoromethane	7.59	113	357206	44.98	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	89.96%	
50) Toluene-d8	10.09	98	1259428	43.52	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	87.04%	
62) 4-Bromofluorobenzene	12.40	95	347668	34.88	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	69.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	16440	8.48	ug/l	# 85
18) Methyl Acetate	4.33	43	29645	4.11	ug/l	98
20) Methylene Chloride	4.55	84	313977	42.60	ug/l	98
43) Isopropyl Acetate	8.17	43	403312	31.96	ug/l	# 83
59) 2-Hexanone	10.77	43	14932	3.15	ug/l	# 58
95) Naphthalene	15.14	128	27660	6.31	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
Data File : VN049980.D
Acq On : 20 Jul 2018 22:51
Operator : MD\SY
Sample : J4057-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LS-SOIL-2

Quant Time: Jul 21 03:22:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

