

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
 Data File : VN049988.D
 Acq On : 21 Jul 2018 4:22
 Operator : MD\SY
 Sample : J4087-53
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ES-PILLAR-3-4

Quant Time: Jul 21 05:41:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	558463	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	923708	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	759765	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	277410	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	404398	52.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.38%	
35) Dibromofluoromethane	7.59	113	351560	45.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.68%	
50) Toluene-d8	10.09	98	1241420	43.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.88%	
62) 4-Bromofluorobenzene	12.40	95	331000	34.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	68.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.41	74	6686	1.71	ug/l	98
16) Acetone	3.82	43	40509	20.38	ug/l	98
18) Methyl Acetate	4.34	43	14692	1.55	ug/l	92
20) Methylene Chloride	4.55	84	1058456	140.05	ug/l	98
25) 2-Butanone	6.85	43	23497	8.27	ug/l	93
29) Tetrahydrofuran	7.22	42	16340	9.13	ug/l	96
95) Naphthalene	15.14	128	30988	6.64	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072018\
Data File : VN049988.D
Acq On : 21 Jul 2018 4:22
Operator : MD\SY
Sample : J4087-53
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ES-PILLAR-3-4

Quant Time: Jul 21 05:41:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

