

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
 Data File : VN049990.D
 Acq On : 21 Jul 2018 5:14
 Operator : MD\SY
 Sample : J4062-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Jul 21 06:31:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	541812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	931809	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	771970	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	288946	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	401880	53.97	ug/l	0.00
Spiked Amount						
			Recovery	=	107.94%	
35) Dibromofluoromethane	7.59	113	352650	45.08	ug/l	0.00
Spiked Amount						
			Recovery	=	90.16%	
50) Toluene-d8	10.09	98	1242094	43.58	ug/l	0.00
Spiked Amount						
			Recovery	=	87.16%	
62) 4-Bromofluorobenzene	12.40	95	339168	34.55	ug/l	0.00
Spiked Amount						
			Recovery	=	69.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	18900	9.80	ug/l	99
18) Methyl Acetate	4.33	43	23388	3.09	ug/l	92
20) Methylene Chloride	4.55	84	83597	11.40	ug/l	95
43) Isopropyl Acetate	8.17	43	235270	18.93	ug/l	97
59) 2-Hexanone	10.77	43	16323	3.50	ug/l #	59
68) m/p-Xylenes	11.62	106	17197	1.44	ug/l	92
95) Naphthalene	15.14	128	43522	7.39	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
Data File : VN049990.D
Acq On : 21 Jul 2018 5:14
Operator : MD\SY
Sample : J4062-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Jul 21 06:31:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

