

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
 Data File : VN049991.D
 Acq On : 21 Jul 2018 5:40
 Operator : MD\SY
 Sample : J4062-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Jul 21 06:32:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	554389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	945783	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	788769	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	294840	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	401941	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
35) Dibromofluoromethane	7.59	113	359341	45.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.52%	
50) Toluene-d8	10.09	98	1257101	43.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.90%	
62) 4-Bromofluorobenzene	12.40	95	345786	34.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	16564	8.40	ug/l #	87
18) Methyl Acetate	4.34	43	23230	2.97	ug/l	97
20) Methylene Chloride	4.55	84	61897	8.25	ug/l	98
43) Isopropyl Acetate	8.17	43	233162	18.48	ug/l	97
59) 2-Hexanone	10.77	43	16175	3.41	ug/l #	58
68) m/p-Xylenes	11.62	106	12780	1.05	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	33023	1.82	ug/l	99
95) Naphthalene	15.14	128	9089	5.08	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
Data File : VN049991.D
Acq On : 21 Jul 2018 5:40
Operator : MD\SY
Sample : J4062-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Jul 21 06:32:54 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

