

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072018\
 Data File : VN049992.D
 Acq On : 21 Jul 2018 6:06
 Operator : MD\SY
 Sample : J4062-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Quant Time: Jul 21 06:34:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	546157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	937170	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	788492	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	320733	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	401198	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
35) Dibromofluoromethane	7.59	113	352341	44.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.56%	
50) Toluene-d8	10.09	98	1257274	43.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.72%	
62) 4-Bromofluorobenzene	12.40	95	351543	35.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.22%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	15312	7.88	ug/l #	80
18) Methyl Acetate	4.34	43	27045	3.66	ug/l	99
20) Methylene Chloride	4.55	84	1369478	185.29	ug/l	99
43) Isopropyl Acetate	8.17	43	253330	20.26	ug/l	94
59) 2-Hexanone	10.77	43	17358	3.70	ug/l #	61
68) m/p-Xylenes	11.62	106	18749	1.54	ug/l	93
84) 1,2,4-Trimethylbenzene	13.04	105	50570	2.56	ug/l	99
95) Naphthalene	15.14	128	406995	28.88	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072018\
Data File : VN049992.D
Acq On : 21 Jul 2018 6:06
Operator : MD\SY
Sample : J4062-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Jul 21 06:34:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

