


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072018\
 Data File : VN049955.D
 Acq On : 20 Jul 2018 11:03
 Operator : MD\SY
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Title : SW846 8260
 Last Update : Thu Jul 19 01:58:39 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3349

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	9944	PASS
75	95	30	60	50.7	25914	PASS
95	95	100	100	100.0	51064	PASS
96	95	5	9	7.1	3641	PASS
173	174	0.00	2	1.8	896	PASS
174	95	50	100	97.5	49784	PASS
175	174	5	9	6.9	3417	PASS
176	174	95	101	96.4	47970	PASS
177	176	5	9	6.7	3192	PASS