

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072020\
 Data File : VN062559.D
 Acq On : 20 Jul 2020 14:05
 Operator : JC/MD
 Sample : L3331-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE-(JULY)

Quant Time: Jul 21 01:25:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071320W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Jul 14 00:52:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	36059	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	187426	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	176340	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.99	65	96141	31.04	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.47%
60) 4-Bromofluorobenzene	12.38	95	78823	26.27	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	87.57%
63) Toluene-d8	10.06	98	240514	28.86	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	96.20%

Target Compounds

					Ovalue
12) Methyl Acetate	4.28	43	6759m	2.427	ug/l
15) Acetone	3.78	58	516875	1300.023	ug/l
30) 2-Butanone	6.80	43	18839	11.173	ug/l #
58) 4-Methyl-2-Pentanone	9.96	43	10883	3.168	ug/l
62) Toluene	10.13	91	272819	27.271	ug/l

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072020\
Data File : VN062559.D
Acq On : 20 Jul 2020 14:05
Operator : JC/MD
Sample : L3331-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE-(JULY)

Quant Time: Jul 21 01:25:00 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071320W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Jul 14 00:52:57 2020
Response via : Initial Calibration

