

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072020\
 Data File : VN062562.D
 Acq On : 20 Jul 2020 15:38
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jul 21 01:32:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071320W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Jul 14 00:52:57 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	94	0.00
2 M	Dichlorodifluoromethane	1.771	1.757	0.8	94	0.00
3 M	Chloromethane	2.557	2.310	9.7	87	0.00
4 M	Vinyl Chloride	2.235	2.127	4.8	91	0.00
5 M	Bromomethane	1.109	0.902	18.7	83	0.00
6 M	Chloroethane	1.100	1.050	4.5	90	0.00
7 M	Trichlorofluoromethane	2.619	2.754	-5.2	97	0.00
8 T	Diethyl Ether	1.259	1.227	2.5	92	0.00
9	1,1,2-Trichlorotrifluoroeth	1.693	1.706	-0.8	96	0.00
10 M	1,1-Dichloroethene	1.728	1.694	2.0	94	0.00
11	Methyl Iodide	2.116	1.319	37.7#	61	0.00
12	Methyl Acetate	2.317	2.164	6.6	90	0.00
13 M	Acrolein	0.346	0.253	26.9	68	0.00
14 M	Acrylonitrile	1.078	0.976	9.5	88	0.00
15 M	Acetone	0.331	0.264	20.2	64	0.00
16 M	Carbon Disulfide	5.630	5.312	5.6	92	0.00
17	Allyl chloride	3.674	3.234	12.0	80	0.00
18 M	Methylene Chloride	2.126	2.091	1.6	92	0.00
19 M	trans-1,2-Dichloroethene	1.912	1.872	2.1	95	0.00
20 T	Diisopropyl ether	7.553	7.467	1.1	93	0.00
21 M	1,1-Dichloroethane	3.993	3.942	1.3	94	0.00
22 M	cis-1,2-Dichloroethene	2.156	2.136	0.9	95	0.00
23 M	tert-Butyl Alcohol	0.361	0.307	15.0	79	0.00
24 M	Methyl tert-Butyl Ether	6.369	6.143	3.5	92	0.00
25 M	Chloroform	3.662	3.701	-1.1	96	0.00
26	Cyclohexane	3.520	3.345	5.0	91	0.00
27 s	1,2-Dichloroethane-d4	2.577	2.531	1.8	92	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
29	1,1-Dichloropropene	0.509	0.531	-4.3	97	0.00
30 M	2-Butanone	0.270	0.243	10.0	77	0.00
31	2,2-Dichloropropane	0.574	0.602	-4.9	95	0.00
32 M	1,1,1-Trichloroethane	0.547	0.580	-6.0	97	0.00
33 M	Carbon Tetrachloride	0.474	0.501	-5.7	97	0.00
34 M	Benzene	1.512	1.570	-3.8	95	0.00
35	Methacrylonitrile	0.237	0.205	13.5	76	0.00
36 M	1,2-Dichloroethane	0.543	0.569	-4.8	96	0.00
37 M	Trichloroethene	0.340	0.358	-5.3	97	0.00
38	Methylcyclohexane	0.571	0.560	1.9	92	0.00
39 M	1,2-Dichloropropane	0.431	0.444	-3.0	94	0.00
40	Dibromomethane	0.248	0.261	-5.2	96	0.00
41 M	Bromodichloromethane	0.526	0.553	-5.1	96	0.00
42 M	Vinyl Acetate	1.121	1.116	0.4	89	0.00
43	Ethyl Acetate	0.543	0.541	0.4	91	0.00
44	Isopropyl Acetate	0.910	0.871	4.3	88	0.00
45 T	1,4-Dioxane	0.006	0.005#	16.7	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072020\
 Data File : VN062562.D
 Acq On : 20 Jul 2020 15:38
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jul 21 01:32:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071320W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Jul 14 00:52:57 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.432	0.406	6.0	86	0.00
47	n-amyl Acetate	0.767	0.721	6.0	89	0.00
48 M	t-1,3-Dichloropropene	0.614	0.620	-1.0	95	0.00
49 T	cis-1,3-Dichloropropene	0.663	0.678	-2.3	95	0.00
50 M	1,1,2-Trichloroethane	0.348	0.368	-5.7	96	0.00
51	Ethyl methacrylate	0.586	0.557	4.9	89	0.00
52	1,3-Dichloropropane	0.639	0.668	-4.5	95	0.00
53 M	Dibromochloromethane	0.379	0.392	-3.4	95	0.00
54 M	1,2-Dibromoethane	0.356	0.365	-2.5	96	0.00
55 M	2-Chloroethyl vinyl ether	0.187	0.177	5.3	90	0.00
56 M	Bromoform	0.259	0.262	-1.2	94	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
58 M	4-Methyl-2-Pentanone	0.584	0.548	6.2	86	0.00
59 M	2-Hexanone	0.429	0.383	10.7	82	0.00
60 S	4-Bromofluorobenzene	0.511	0.510	0.2	97	0.00
61 M	Tetrachloroethene	0.307	0.325	-5.9	100	0.00
62 M	Toluene	1.702	1.725	-1.4	95	0.00
63 S	Toluene-d8	1.418	1.406	0.8	93	0.00
64 M	Chlorobenzene	1.000	1.009	-0.9	98	0.00
65	1,1,1,2-Tetrachloroethane	0.370	0.377	-1.9	97	0.00
66 M	Ethyl Benzene	1.885	1.873	0.6	95	0.00
67 M	m/p-Xylenes	0.684	0.690	-0.9	98	0.00
68 M	o-Xylene	0.656	0.646	1.5	95	0.00
69 M	Styrene	1.114	1.102	1.1	96	0.00
70	Isopropylbenzene	1.762	1.768	-0.3	97	0.00
71 M	1,1,2,2-Tetrachloroethane	0.570	0.559	1.9	93	0.00
72	1,2,3-Trichloropropane	0.554	0.520	6.1	89	0.00
73	Bromobenzene	0.408	0.410	-0.5	98	0.00
74	n-propylbenzene	2.095	2.136	-2.0	100	0.00
75	2-Chlorotoluene	1.261	1.275	-1.1	98	0.00
76	1,3,5-Trimethylbenzene	1.502	1.502	0.0	99	0.00
77	t-1,4-Dichloro-2-butene	0.207	0.180	13.0	86	0.00
78	4-Chlorotoluene	1.287	1.318	-2.4	100	0.00
79	tert-butylbenzene	1.243	1.220	1.9	95	0.00
80	1,2,4-Trimethylbenzene	1.477	1.446	2.1	95	0.00
81	sec-Butylbenzene	1.670	1.661	0.5	98	0.00
82	p-Isopropyltoluene	1.474	1.463	0.7	99	0.00
83 M	1,3-Dichlorobenzene	0.709	0.736	-3.8	103	0.00
84 M	1,4-Dichlorobenzene	0.699	0.726	-3.9	105	0.00
85	n-Butylbenzene	1.264	1.191	5.8	100	0.00
86 T	Hexachloroethane	0.284	0.270	4.9	94	0.00
87 M	1,2-Dichlorobenzene	0.682	0.706	-3.5	102	0.00
88	1,2-Dibromo-3-Chloropropane	0.108	0.100	7.4	89	0.00
89	1,2,4-Trichlorobenzene	0.339	0.345	-1.8	108	0.00
90	Hexachlorobutadiene	0.188	0.203	-8.0	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072020\
Data File : VN062562.D
Acq On : 20 Jul 2020 15:38
Operator : JC/MD
Sample : VSTDCCC020
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC020

Quant Time: Jul 21 01:32:43 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071320W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Jul 14 00:52:57 2020
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	0.920	0.784	14.8	91	0.00
92	1,2,3-Trichlorobenzene	0.342	0.343	-0.3	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0