

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
 Data File : VN067812.D
 Acq On : 20 Jul 2021 12:45
 Operator : JC/MD
 Sample : VN0720MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0720MBL01

Quant Time: Jul 20 15:33:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	344945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	637262	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	609284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	209031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	260985	50.063	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.120%	
35) Dibromofluoromethane	8.024	113	195589	50.800	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.600%	
50) Toluene-d8	10.443	98	812440	50.412	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.820%	
62) 4-Bromofluorobenzene	12.733	95	259300	45.547	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 91.100%	
Target Compounds						Qvalue
95) Naphthalene	15.512	128	20430	4.888	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
Data File : VN067812.D
Acq On : 20 Jul 2021 12:45
Operator : JC/MD
Sample : VN0720MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0720MBL01

Quant Time: Jul 20 15:33:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

