

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
 Data File : VN067814.D
 Acq On : 20 Jul 2021 13:42
 Operator : JC/MD
 Sample : M3065-01
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-071521

Quant Time: Jul 20 15:34:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.077	168	313133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	579905	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	569583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	234545	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.431	65	238193	50.332	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.660%
35) Dibromofluoromethane	8.016	113	172530	49.243	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.480%
50) Toluene-d8	10.440	98	760901	51.884	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.760%
62) 4-Bromofluorobenzene	12.736	95	259278	50.048	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
Data File : VN067814.D
Acq On : 20 Jul 2021 13:42
Operator : JC/MD
Sample : M3065-01
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-071521

Quant Time: Jul 20 15:34:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

