

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
 Data File : VN067820.D
 Acq On : 20 Jul 2021 16:23
 Operator : JC/MD
 Sample : PB137795ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137795ZHE#01

Quant Time: Jul 21 05:08:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	333493	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	634420	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	649566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	229930	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	262032	51.989	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	103.980%	
35) Dibromofluoromethane	8.024	113	191993	50.089	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	100.180%	
50) Toluene-d8	10.443	98	820615	51.148	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	102.300%	
62) 4-Bromofluorobenzene	12.734	95	282480	49.841	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	99.680%	
Target Compounds						Qvalue
16) Acetone	4.293	43	25204	15.892	ug/l	# 87
20) Methylene Chloride	5.093	84	27401	6.800	ug/l	90
43) Isopropyl Acetate	8.566	43	41250	4.107	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
Data File : VN067820.D
Acq On : 20 Jul 2021 16:23
Operator : JC/MD
Sample : PB137795ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137795ZHE#01

Quant Time: Jul 21 05:08:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

