

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
 Data File : VN067823.D
 Acq On : 20 Jul 2021 17:36
 Operator : JC/MD
 Sample : PB137795ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137795ZHE#04

Quant Time: Jul 21 05:09:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	310761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	601579	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	619921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	234789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	245924	52.363	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	104.720%	
35) Dibromofluoromethane	8.024	113	180394	49.632	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	99.260%	
50) Toluene-d8	10.446	98	782310	51.422	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	102.840%	
62) 4-Bromofluorobenzene	12.734	95	270252	50.286	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	100.580%	
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	24483	16.567	ug/l #	83
18) Methyl Acetate	4.870	43	6621	1.595	ug/l	96
20) Methylene Chloride	5.095	84	23038	6.066	ug/l	86
43) Isopropyl Acetate	8.566	43	44023	4.622	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
Data File : VN067823.D
Acq On : 20 Jul 2021 17:36
Operator : JC/MD
Sample : PB137795ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137795ZHE#04

Quant Time: Jul 21 05:09:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

