

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
 Data File : VN067826.D
 Acq On : 20 Jul 2021 18:50
 Operator : JC/MD
 Sample : M3089-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-N2-B-BOTTOM-(12)

Quant Time: Jul 21 05:11:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	308799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	590804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	602094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	227544	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	241188	51.681	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.360%
35) Dibromofluoromethane	8.024	113	177199	49.643	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.280%
50) Toluene-d8	10.446	98	762027	51.002	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.000%
62) 4-Bromofluorobenzene	12.733	95	259520	49.170	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.340%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	50418	34.333	ug/l	96
18) Methyl Acetate	4.870	43	8207	1.990	ug/l #	78
20) Methylene Chloride	5.095	84	32996	9.055	ug/l	93
43) Isopropyl Acetate	8.563	43	30189	3.228	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072021\
Data File : VN067826.D
Acq On : 20 Jul 2021 18:50
Operator : JC/MD
Sample : M3089-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-N2-B-BOTTOM-(12)

Quant Time: Jul 21 05:11:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

