

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072022\
 Data File : VN073525.D
 Acq On : 20 Jul 2022 16:16
 Operator : JC\MD
 Sample : VN0720WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0720WBL01

Quant Time: Jul 21 08:24:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	224239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	374059	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	352550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	167633	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	153010	47.267	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.540%
35) Dibromofluoromethane	7.951	113	126827	50.962	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.920%
50) Toluene-d8	10.380	98	407658	46.121	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.240%
62) 4-Bromofluorobenzene	12.674	95	164006	48.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072022\
Data File : VN073525.D
Acq On : 20 Jul 2022 16:16
Operator : JC\MD
Sample : VN0720WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0720WBL01

Quant Time: Jul 21 08:24:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

