

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072022\
 Data File : VN073539.D
 Acq On : 20 Jul 2022 22:13
 Operator : JC\MD
 Sample : N3779-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 070622-E

Quant Time: Jul 21 08:27:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	297854	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	487849	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	465244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	261306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	185246	43.082	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.160%		
35) Dibromofluoromethane	7.951	113	157432	48.504	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.000%		
50) Toluene-d8	10.381	98	522905	45.361	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.720%		
62) 4-Bromofluorobenzene	12.669	95	241706	54.560	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	109.120%		
Target Compounds					Qvalue	
16) Acetone	4.222	43	204440	84.704	ug/l	96
17) Carbon Disulfide	4.457	76	24347	2.678	ug/l #	90
25) 2-Butanone	7.269	43	26514	6.900	ug/l #	73
30) Chloroform	7.751	83	28439	3.959	ug/l	93
40) Benzene	8.392	78	287382	18.646	ug/l	100
47) Bromodichloromethane	9.698	83	11438	2.214	ug/l	92
52) Toluene	10.445	92	192760	20.237	ug/l	99
64) Tetrachloroethene	10.922	164	2549	0.671	ug/l #	84
67) Ethyl Benzene	11.786	91	49788	2.606	ug/l	99
68) m/p-Xylenes	11.892	106	122828	17.037	ug/l	99
69) o-Xylene	12.222	106	88975	12.090	ug/l	94
78) n-propylbenzene	12.863	91	9353	0.373	ug/l	97
80) 1,3,5-Trimethylbenzene	12.998	105	41683	2.266	ug/l	97
84) 1,2,4-Trimethylbenzene	13.310	105	154579	8.418	ug/l	94
95) Naphthalene	15.433	128	4793805	238.034	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072022\
Data File : VN073539.D
Acq On : 20 Jul 2022 22:13
Operator : JC\MD
Sample : N3779-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
070622-E

Quant Time: Jul 21 08:27:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

