

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072023\
 Data File : VN078591.D
 Acq On : 20 Jul 2023 13:45
 Operator : JC\MD
 Sample : 03670-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 N48909

Quant Time: Jul 21 06:11:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	436715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	762734	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	706649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	243323	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	281891	48.148	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.300%
35) Dibromofluoromethane	8.177	113	285186	55.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.160%
50) Toluene-d8	10.571	98	911176	46.584	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.160%
62) 4-Bromofluorobenzene	12.853	95	305980	50.787	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.580%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	34295	18.805	ug/l	90
25) 2-Butanone	7.506	43	20861	7.267	ug/l	99
29) Tetrahydrofuran	7.847	42	48951	25.986	ug/l #	82
65) Chlorobenzene	11.900	112	5764	0.365	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072023\
Data File : VN078591.D
Acq On : 20 Jul 2023 13:45
Operator : JC\MD
Sample : 03670-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
N48909

Quant Time: Jul 21 06:11:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

