

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072023\
 Data File : VN078600.D
 Acq On : 20 Jul 2023 17:58
 Operator : JC\MD
 Sample : 03632-06ME
 Misc : 12.08g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-09-(3-5)ME

Quant Time: Jul 21 06:19:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	409919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	719772	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	718226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	289939	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	262721	47.807	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.620%
35) Dibromofluoromethane	8.177	113	258093	52.823	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.640%
50) Toluene-d8	10.571	98	872677	47.279	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.560%
62) 4-Bromofluorobenzene	12.853	95	333492	58.657	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072023\
Data File : VN078600.D
Acq On : 20 Jul 2023 17:58
Operator : JC\MD
Sample : 03632-06ME
Misc : 12.08g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-09-(3-5)ME

Quant Time: Jul 21 06:19:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

