

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072120\
 Data File : VN062575.D
 Acq On : 21 Jul 2020 16:51
 Operator : JC/MD
 Sample : L3342-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-6

Quant Time: Jul 22 06:27:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	165583	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	288353	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	266624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	92529	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	127719	49.65	ug/l	0.00
Spiked Amount			Recovery	=	99.30%	
35) Dibromofluoromethane	7.55	113	94056	49.66	ug/l	0.00
Spiked Amount			Recovery	=	99.32%	
50) Toluene-d8	10.06	98	365013	48.30	ug/l	0.00
Spiked Amount			Recovery	=	96.60%	
62) 4-Bromofluorobenzene	12.38	95	124667	45.12	ug/l	0.00
Spiked Amount			Recovery	=	90.24%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072120\
Data File : VN062575.D
Acq On : 21 Jul 2020 16:51
Operator : JC/MD
Sample : L3342-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-6

Quant Time: Jul 22 06:27:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

